



CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.
One Micromeritics Drive
Norcross, Georgia 30093-1877 U.S.A.
ISO 9001:2000 Certified

VERIFICATION

Model Number: Elzone II 5390

Type of Equipment: Particle Size Analyzer

Standards of Conformity: EN 61326:1997+A1:1998+A2:2001 – Electrical equipment for measurement, control and Laboratory use – EMC requirements

- | | |
|----------------------------|---|
| EN 61000-4-2:95 | Electronic Discharge Immunity |
| EN 61000-4-3:96 | RF Electromagnetic Field |
| EN 61000-4-4:95 | Electrical Fast Transient/Burst |
| EN 61000-4-5:95 | Surge Immunity |
| EN 61000-4-6:96 | RF Conducted Disturbances |
| EN 61000-4-11:1994 | Voltage Dips, Interruptions, and Variations |
| EN 61326 (CE) | Radiated Emissions |
| EN 61326 (CE) | Conducted Emissions |
| EN 61000-3-2:1995+A14:1998 | Harmonic Current Emissions |
| EN 61000-3-3:1995 | Voltage Fluctuations |

EN 61010-1 Part 1: General Safety Requirements

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.

Stefan Koch
Signature

July 11 - 2007
Date

Stefan Koch
Print name

Director of Engineering
Position