



CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.
4356 Communications Dr.
Norcross, Georgia 30093-2901 U.S.A.
ISO 9001:2000 Certified

VERIFICATION

Model Number: TriStar II 3020

Type of Equipment: Surface Area and Porosity Analyzer


Applicable Directive: **2004/108/EEC** European Union (EU) Emissions and Immunity
73/23/EEC Low Voltage Directive European Union (EU) Safety

Standards of Conformity:

| | |
|--|---|
| EN 61326: 1997 +A1: 1998 +A2:2001 | |
| IEC 61000-4-2 +A1+A2:2000 | Electrostatic Discharge |
| EN 61326 CISPR 11:2003 (RE) | Radiated Emissions |
| EN 61326 CISPR 11:2003 (CE) | Conducted Emissions |
| IEC 61000-4-3:2002 | Radiated RF Electromagnetic Immunity |
| IEC 61000-3-2:2000 | Harmonic Current Emissions |
| IEC 61000-3-3 +A1:2001 | Voltage Fluctuations and Flicker |
| IEC 61000-4-4:2004 | Electrical Fast Transient/Burst |
| IEC 61000-4-5 +A1:2000 | Surge Immunity |
| IEC 61000-4-6:2003 | Conducted Immunity |
| IEC 61000-4-11:2004 | Voltage Dips, Interruptions, and Variations |
| EN 61010-1 Part 1: | General Safety Requirements |

European Contact Address: Micromeritics Germany GmbH, Avantis Science Park, Rutherford 108, D-52072 Aachen
Tel: +49 (0) 241 189 446 0, Fax: +49 (0) 241 189 446 11
email:micromeritics.de@micromeritics.com

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.



Signature

3 - 28 - 2011

Date

Stefan Koch

Print name

Director of Engineering.

Position